








Minutes for meeting for Pre-bid conference for tender no. GEM/2025/B/5962110 for Video Spectral Comparator

Sr.no.	Firm Name	Reference Clause of tender	Description of query	ISP reply
1	M/s Perceptive Biotech Pvt. Ltd.	Specifications Sr No 2 – Spectral range 350-1100nm with a 80 X or more optical range	Spectral range 350-1100nm with a 480 X or more optical range	The request of firm shall be reviewed. If needed, suitable corrigendum will be issued
	M/s Perceptive Biotech Pvt. Ltd.	Specification Sr No 4 – Digital zoom up to 8X giving image magnification up to 1000x or better (dependent on display monitor)	Digital zoom of 10X or more giving image magnification of 5000x or better on 32" 4K monitor	The request of firm shall be reviewed. If needed, suitable corrigendum will be issued
	M/s Perceptive Biotech Pvt. Ltd.	Specifications Sr No 11. iii) Two IR/visible oblique lights, Suitable Number of white and IR LEDs with programmable variable angle selection	Two IR/visible oblique lights, Suitable Number of white and IR LEDs with manual or programmable variable angle selection	The request of firm shall be reviewed. If needed, suitable corrigendum will be issued
2	M/s Vajra BioMetric Pvt. Ltd.	Experience criteria-The bidder should have manufactured and supplied at least one number of "Video Spectral Comparator" with the same or higher specification in last five years ending 31.03.2024. At least one number of the product offered for supply should be in successful operation for at least one year on the date of bid opening.	The bidder should have manufactured and supplied at least one number of "Video Spectral Comparator "with the same or higher specification in the last five years ending 31.03.2024. At least one number of the product offered for supply should be in successful operation for at least six (6) months on the date of bid opening. "	Clarified. The criteria shall be as per GEM bid terms and conditions.
	M/s Vajra BioMetric Pvt. Ltd.	Specifications Sr No 10. ii)Minimum 16 long pass filters of wavelength ranging from 515 to 925 nm or better	Minimum 16 long pass filters of wavelength ranging from 515 to 920 nm or better	The request of firm shall be reviewed. If needed, suitable corrigendum will be issued

Sr.no.	Firm Name	Reference Clause of tender	Description of query	ISP reply
	M/s Vajra BioMetric Pvt. Ltd.	Specifications Sr No 15. i)Integral grating spectrometer for measurement of absorption, reflectance, transmission and fluorescence spectra from 350 nm to 1000nm with a resolution of 2nm or better.	Integral grating spectrometer for measurement of absorption, reflectance, transmission and fluorescence spectra from 350 nm to 1000nm with a resolution of 2.5nm or better.	The request of firm shall be reviewed. If needed, suitable corrigendum will be issued
	M/s Vajra BioMetric Pvt. Ltd.	OPTIONAL REQUIREMENTS	Regarding defining optional requirements	The request of firm shall be reviewed. If needed, suitable corrigendum will be issued

Sr.no.	Name of ISP officials attending the pre- bid conference	Signature
1	Shri. Diganta Kumar Deka, AGM(M)	 03.03.2025
2	Shri. Nilesh Jadhav, JGM(TO)	 03/03/2025
3	Shri. Vikas Kumar, M(MM)	 25/3
4	Shri. K Jayprakash Reddy, DM (TO)	 03/03/25
5	Shri. B. P. Rajan, Sr. Supervisor(TO)	 25/03/25

Sr.no.	Name of firm attending the pre- bid conference	Name of Firm's representative attending the pre- bid conference	Signature
1	M/s Perceptive Biotech Pvt. Ltd.	Shri. Ketan Kambli	
2	M/s Vajra BioMetric Pvt. Ltd	Shri. L. Praveen Chand	 21/3/20
3	M/s Vajra BioMetric Pvt. Ltd	Shri. Nikhil N Chippa	